

Organic matter removal for continuous flow isotope ratio mass spectrometry analysis of carbon and oxygen isotope compositions of calcite or dolomite in organic-rich samples

Carine Chaduteau, Magali Ader, Oanez Lebeau, Guillaume Landais, Vincent Busigny

▶ To cite this version:

Carine Chaduteau, Magali Ader, Oanez Lebeau, Guillaume Landais, Vincent Busigny. Organic matter removal for continuous flow isotope ratio mass spectrometry analysis of carbon and oxygen isotope compositions of calcite or dolomite in organic-rich samples. Limnology and Oceanography: Methods, 2021, 19 (8), pp.523-539. 10.1002/lom3.10442. hal-03822129

HAL Id: hal-03822129 https://u-paris.hal.science/hal-03822129

Submitted on 20 Oct 2022

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers. L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.



LIMNOLOGY and OCEANOGRAPHY: METHODS



Limnol. Oceanogr.: Methods 19, 2021, 523–539
© 2021 The Authors. Limnology and Oceanography: Methods published by
Wiley Periodicals LLC on behalf of Association for the Sciences of
Limnology and Oceanography.
doi: 10.1002/lom3.10442

Organic matter removal for continuous flow isotope ratio mass spectrometry analysis of carbon and oxygen isotope compositions of calcite or dolomite in organic-rich samples

Carine Chaduteau, ¹ Magali Ader, ¹* Oanez Lebeau, ² Guillaume Landais, ¹ Vincent Busigny ^{1,3}

Abstract

 δ^{13} C and δ^{18} O values of carbonate in organic-rich samples such as soils, biofilms, and lake sediments have been scarcely used so far in paleo-environmental, biomineralization, or diagenetic studies because organic matter in high proportions is suspected to alter carbonate isotope analysis. Yet, with the improvement of analytical capabilities and in particular the use of CF-IRMS, this may not be an issue anymore. To evaluate this, δ^{13} C and δ^{18} O values of calcite or dolomite mixed in various proportions with yeast (used here as a model for immature organic matter) were measured. The results indicate that measurements of calcite or dolomite δ^{13} C and δ^{18} O by CF-IRMS require organic matter removal only when its weight proportion exceeds that of carbonates. Yeast generated an unidentified molecule during phosphoric acid digestion at 25°C (none at 80°C), which shifted the carbonate δ^{18} O values for yeast proportions higher than 50%. It also generated CO₂, but with a noticeable shift in δ^{13} C and δ^{18} O values only at 80°C (none at 25°C) and for yeast proportions higher than 75%. Three methods of organic matter removal were tested: the well-established NaOCl and H₂O₂ wet treatments as well as low temperature plasma ashing (LTA) so far seldom examined. LTA removed the isotope shifts, although imperfectly for yeast proportions of 95%, without carbonate dissolution. NaOCl efficiently removed the isotope shifts in all cases but dissolved part of the calcite and dolomite. H₂O₂ treatment caused severe dissolution of calcite and dolomite while worsening the δ^{13} C and δ^{18} O shifts and should thus be avoided.

Carbon and oxygen isotope analyses of calcite ($CaCO_3$) and dolomite ($CaMg[CO_3]_2$) in sediments, soils, shells, speleothems, microbialites, and sedimentary rocks are routinely performed to reconstruct paleo-environmental changes or diagenetic history. In most cases, these analyses do not require pretreatment to remove organic matter because carbonates are in large excess relative to organic matter (e.g., Wierzbowski 2007). Pretreatments remain however sometimes necessary, especially when samples contain immature organic matter. CO_2 and/or other molecular gaseous species may be produced during the reaction of organic matter with orthophosphoric acid (H_3PO_4) (Bowen 1966, 1991; Epstein et al. 1951, 1953; Falster et al. 2018; Oehlerich

Additional Supporting Information may be found in the online version of this article.

This is an open access article under the terms of the Creative Commons Attribution License, which permits use, distribution and reproduction in any medium, provided the original work is properly cited.

et al. 2013; Weber et al. 1976) and their incomplete separation prior to gas introduction in the mass spectrometer may affect C and O isotope measurements. Various treatments for organic matter removal have thus been tested (see reviews in Grottoli et al. 2005; Keatings et al. 2006; De Groot 2011; Falster et al. 2018), for instance in the view of defining standard procedures for C and O isotope analyses of foraminifera (e.g., Feldmeijer et al. 2013), ostracod (e.g., Keatings et al. 2006), corals, and/or coccolithophores (e.g., Stevenson et al. 2014). However, these treatments were tested only for relatively low organic matter to carbonate ratios (<15%, Wierzbowski 2007), typical of most geological samples, except for H₂O₂ treatment, tested recently by Falster et al. (2018) for extremely high ratios. Although they are not very common in sedimentary rocks, high ratios can be found in soils, in biofilms or more generally in organic-rich and/or carbonate-poor marine or lacustrine sediments. In addition, the effect of organic matter removal treatments on carbonate isotope compositions has mostly been tested using vacuum CO₂ extraction and cryogenic purification systems such as vacuum lines (e.g., Wierzbowski 2007) or Kiel

¹Université de Paris, Institut de Physique du Globe de Paris, CNRS, Paris, France

²IUEM, Laboratoire Domaines Océaniques, Université de Bretagne Occidentale, Plouzane, France

³Institut Universitaire de France, Paris, France

^{*}Correspondence: ader@ipqp.fr

devices (e.g., Nagtegaal et al. 2012). Many laboratories now use continuous flow methods with a gas chromatographic column separation (CF-IRMS), which purifies CO₂ from other gases by gas chromatography rather than by cryogenic separation and may incompletely separate CO₂ from other contaminant gaseous species possibly generated during phosphoric acid attack. Last, all the recent CF-IRMS studies used a phosphoric acid digestion temperature ranging from 72°C to 90°C (Falster et al. 2018; Feldmeijer et al. 2013; Oehlerich et al. 2013; Keatings et al. 2006), i.e., quite high compared to the temperature of 25°C used for calcite in most earlier studies. Such high temperatures might increase the reactivity of organic matter with phosphoric acid.

In the present work, δ^{13} C and δ^{18} O values of calcite or dolomite mixed in various proportions with yeast (used here as a model for immature organic matter) were digested respectively at 25°C and 80°C and measured by CF-IRMS in order to evaluate the need for organic matter removal. The proportions of yeast included the extreme values of 75%, 90%, and 95% hardly ever investigated so far to our knowledge. Three widely used organic matter removal treatments were also tested and compared: wet oxidation with NaOCl (e.g., Charef and Sheppard 1984; Durazzi 1977; Emiliani 1966; Grottoli et al. 2005; Irwin et al. 1977; Land et al. 1975); wet oxidation with H₂O₂ (e.g., Boiseau and Juillet-Leclerc 1997; D'Eugenio and Leone 1989; D'Hondt and Lindinger 1994; McConnaughey 1989) also often used before Δ_{47} measurements (e.g., Grauel et al. 2013; Katz et al. 2017; Peral et al. 2018; Tripati et al. 2010) and dry oxidation using low-temperature oxygenplasma ashing (LTA) (Conan and Brummer 2000; Erez and Honjo 1981; Goreau 1977; Hendry and Kalin 1997; Loncaric et al. 2007; Swart and Coleman 1980;). Roasting was not investigated here as several previous studies demonstrated that this treatment yields unsatisfactory results (e.g., Boiseau and Juillet-Leclerc 1997; Erez and Honjo 1981; Gaffey et al. 1991; Grottoli et al. 2005; Guiguer et al. 2003; Leone et al., 2000; Wierzbowski 2007). Phosphoric acid digestion was performed at 25°C for calcite samples and 80°C for dolomite samples, allowing to test the impact of temperature on the genesis of CO₂ and/or other contaminants from immature organic matter before and after treatment.

Materials and methods

Sample preparation

Two series of samples were prepared from variable proportions of either yeast and calcite or yeast and dolomite. The calcite (sample RENNES 0, labeled R_0) is a synthetic powder from Prolabo (99.5%), used in our laboratory as an internal isotope standard. The dolomite (sample INYO) is a natural sample of dolomitic marble from the Inyo mountains in California (USA). The dolomite was crushed to fine grain (<50 μ m) using an agate mortar and thoroughly homogenized. Yeast was purchased in a 500 g box as dry balls of 1 mm diameter (cf. LEVURE-SECHE at "L'atelier de la patisserie"). It was also crushed to a fine powder in an agate mortar. Yeast (i.e., *Saccharomyces cerevisia*) was selected to represent immature organic matter because it

contains all the molecular constituents of living organisms. It is assumed here to represent organic matter at the initial stages of early diagenesis. In addition, it can be easily purchased and preserved under dried and cooled conditions without any evolution of its C isotope composition. Variable amounts of carbonate and yeast were weighted and mixed directly in glass vials to obtain homogeneous synthetic samples with a yeast proportion of 10, 25, 50, 75, 90, and 95 wt%.

Treatments for organic matter removal

Wet oxidation by H₂O₂ or NaOCl

The wet oxidation reactions took place in a chemical fume hood at room temperature (21 \pm 2°C). Aliquots of synthetic sample powders were placed into glass beakers with reagent grade hydrogen peroxide 30% (H₂O₂, pH = 3.8) (VWR Prolabo, AnalaR Normapur) or sodium hypochlorite (NaOCl, pH = 12) with 3.5% of active chlorine (VWR Prolabo, GPR Rectapur). Sample masses, reagent volumes, and reaction times are given in the Table S1. Two protocols were tested with H₂O₂ solution. The H₂O₂-A protocol corresponds to a maximum yeast proportion of 16.9 g per liter of H₂O₂ (i.e., 40 ml of H₂O₂ for 750 mg of yeast) and a reaction time of 24 h. The H₂O₂-B protocol corresponds to a maximum yeast proportion of 2.9 g per liter of H₂O₂ (i.e., 200 ml of H₂O₂ for 600 mg of yeast) and a reaction time of 72 h. Two protocols were tested for NaOCl solutions. The NaOCl-A protocol lasted 24 h with a maximum yeast proportion of 6.8 g per liter of NaOCl (i.e., 100 ml of NaOCl for 750 mg of yeast). The NaOCl-B protocol lasted 48 h with a maximum yeast proportion of 4.8 g per liter of NaOCl (i.e., 120 ml of NaOCl for 600 mg of yeast). After reaction, samples were centrifuged and rinsed five times with MilliQ water. Finally the residual powders were dried in an oven at 40°C for at least 2 d.

Dry oxidation by low temperature oxygen-plasma ashing

Aliquots of synthetic samples were spread in glass petri dishes and covered with riddled aluminum foil. Sample weight was adjusted so that it contained 15 mg of carbonates, a comfortable amount for subsequent isotopic analyses. The amount of yeast to be removed was thus comprised between 0 and 285 mg. The petri dishes were placed in a low-temperature oxygen-plasma asher (Polaron PT7160 RF) and evacuated until 10⁻¹ mbar. Oxygen was then injected in the asher, and an electric discharge lighted the plasma. This plasma was made of free oxygen radicals, which oxidized organic matter to CO₂. The treatment was initially applied for 1–8 h depending on the amount of yeast and the powder thickness in the petri dishes. A few samples were treated again for up to seven additional hours.

Carbonate quantification and carbon and oxygen isotope analyses

Carbonate contents as well as carbon and oxygen isotope compositions of treated and untreated samples were measured by CF-IRMS. Both an Analytical Precision 2003 (AP2003) and a Gas Bench coupled with a Delta plus XP were used in this study.

The weighed samples were placed in 12 ml borosilicate glass vials (Labco exetainer), sealed with a rubber septum and flushed with helium 5.6 (>99.9996%). About 0.1 ml of 100% $\rm H_3PO_4$ was manually injected using a syringe. Phosphoric acid was prepared without potassium dichromate ($\rm K_2Cr_2O_7$), to avoid any production of $\rm CO_2$ by yeast oxidation with $\rm K_2Cr_2O_7$ (Sarkar et al. 1990). The samples were reacted for 4 h at 25°C for calcite and 2 h at 80°C for dolomite.

For measurements using the AP2003, the gas was sampled automatically from vial headspace by a Gilson 22X autosampler equipped with a capillary needle. The gas was transferred by helium continuous flow and passed through a Nafion membrane to remove $\rm H_2O$. It was then introduced in a chromatographic column (stainless column 0.3 m \times 1/8" \times 2 mm, packed with Hayesep Q 60/80 mesh, Chrompack) in order to separate $\rm CO_2$ from other gas components (e.g., $\rm N_2$, $\rm O_2$, Ar). The $\rm CO_2$ was finally transferred to the mass spectrometer. The $\rm \delta^{13}C_{\rm CO2}$ and $\rm \delta^{18}O_{\rm CO2}$ values of the sample $\rm CO_2$ were calculated relative to a $\rm CO_2$ reference gas of 5.2 purity grade (>99.9992%) injected between each sample peak. For each vial, four successive analyses were carried out and averaged to provide the $\rm \delta$ values.

The Gas Bench coupled with a Delta plus XP was operated according to the same principle as the AP2003 with only minor differences. The gas from the vials was sampled automatically via a CTC Combi-Pal autosampler. The gas chromatography column was a fused silica Poraplot Q column (25 m \times 0.32 mm) heated to 70°C. Each IRMS measurement cycle started by five injections of CO2 reference gas, followed by six aliquots of each sample.

For both mass spectrometers, measurements were performed in batches, which included several sets of internal standards. Standards of pure carbonate (RENNES 0 or INYO) of identical weight were placed every six samples to check the absence of drifts. A set of six standards of various weights of RENNES 0 or INYO was used to calibrate the CO2 signal intensity relative to the carbonate amount, allowing the quantification of the carbonate content of each sample. The external precision (2σ) for the carbonate content estimated from replicate analyses of standards was $\pm 10\%$ relative to the measured value. A set of three calcite internal standards with various isotopic compositions was used to calibrate the $\delta^{13}C$ and $\delta^{18}O$ values relative to PDB (Pee Dee Belemnite). We report the δ^{13} C and $\delta^{18}O$ values of the CO_2 produced by phosphoric acid attack ($\delta^{18}O_{CO2}$ and $\delta^{13}C_{CO2}$) rather than those of carbonate. This allows us to perform mass and isotope balance calculation showing that part of the sample's CO2 was derived from the partial oxidation of yeast (see first section of the discussion). The external precision (20) estimated from replicate analyses on RENNES 0 (n=32) was $\pm 0.14\%$ 0 and $\pm 0.10\%$ for $\delta^{13}C_{CO2}$ and $\delta^{18}O_{CO2}$ respectively. Each synthetic sample was measured at least four times (i.e., twice in two different batches).

The $\delta^{13}C_{CO2}$ and $\delta^{18}O_{CO2}$ values of calcite internal standards, RENNES 0 and INYO, were measured using the classical method of McCrea (1950) in order to calibrate them against the NBS-19 and IAEA-CO-1 international standards. The McCrea method was applied as follows. The sample was reacted with phosphoric acid under vacuum at 25°C for calcite for 4 h and at 80°C for 2 h for dolomite. The evolved CO₂ was then cryogenically purified in a vacuum line using a temperature controlled liquid-N₂ trap set at -130° C and then manometrically quantified. Its isotope composition was analyzed with Delta Plus XP or MAT253 mass spectrometers in Dual Inlet mode. The $\delta^{13}C_{CO2}$ and $\delta^{18}O_{CO2}$ values of a few untreated calcite-yeast mixture samples were also analyzed using this method in order to compare cryogenic purification on vacuum line with gas chromatography separation.

The δ^{13} C value of yeast was measured using sealed-tube combustion with copper oxide (i.e., Dumas combustion), offline vacuum cryogenic purification and quantification of CO₂, and isotope measurement on a Delta Plus XP mass spectrometer (for more details on the method, *see* Ader et al., 1998).

Results

The main goal of the present work is to assess the difference in $\delta^{13}C_{CO2}$ and $\delta^{18}O_{CO2}$ measured for synthetic mixtures of organic carbon and carbonate from values expected in pure carbonate standards (i.e., RENNES 0 or INYO). The results for calcite and dolomite are reported in Tables 1 and 2, respectively. The isotope deviations of the sample mixtures from the pure standard carbonates (also called isotope shifts) are expressed as Δ^{13} C and Δ^{18} O, with Δ^{13} C = δ^{13} C_{CO2-sample} – δ^{13} C CO2-standard and Δ^{18} O = δ^{18} O CO2-sample - δ^{18} O CO2-standard. The deviation is considered as significant when Δ^{13} C and Δ^{18} O values are higher than 0.2‰, i.e., slightly higher than the 2σ precision for $\delta^{13}C$ and $\delta^{18}O$. Additionally, in order to evaluate the potential contribution of CO₂ produced by yeast during H_3PO_4 digestion, the $\delta^{13}C_{CO2}$ and $\delta^{18}O_{CO2}$ (vs. PDB) values of the evolved CO2 produced by pure yeast H3PO4 digestion were also determined (data in Tables 1 and 2).

Untreated samples

Calcite and calcite-yeast mixtures—H₃PO₄ digestion at 25°C

The $\delta^{13}C_{CO2}$ and $\delta^{18}O_{CO2}$ values for pure calcite R0 obtained by CF-IRMS are $-9.82\% \pm 0.14\%$ and $-9.12\% \pm 0.10\%$ respectively (2σ , n=32) and are identical to the values obtained by the classical method of McCrea (1950) which uses cryogenic purification rather than gas chromatographic purification (Table 1). All synthetic mixtures have $\delta^{13}C_{CO2}$ values similar to pure calcite R0, even for high yeast content (Fig. 1A). In contrast, their $\delta^{18}O_{CO2}$ values deviate from that of R0 when yeast proportion is $\geq 50\%$. They increase from -8.86% to -6.78% ($\Delta^{18}O$ from 0.26% to 2.34%) for yeast proportions from 50% to 95% (Fig. 1B). In addition, the

15415856, 2021, 8, Downloaded from https://aslopubs.onlinelibrary.wiley.com/doi/10.1002/lom3.10442 by Cochrane France, Wiley Online Library on [20/10/2022]. See the Terms and Conditions (https://onlinelibrary.wiley.com/erres-ad-conditions) on Wiley Online Library for rules of use; OA articles are governed by the applicable Certaive Commons License

Table 1. $\delta^{13}C_{CO2}$ and $\delta^{18}O_{CO2}$ values of pure calcite, synthetic samples and pure yeast. $\Delta^{13}C$ (= $\delta^{13}C_{CO2-sample} - \delta^{13}C_{CO2-standard}$) and $\Delta^{18}O$ (= $\delta^{18}O_{CO2-sample}$ $\delta^{18}O_{CO2-standard}$) correspond to offsets from the isotopic values of pure calcite.

ite Y °	t Method	(1/~/			ָ										
0 1 2		(g/r)	E	ΑЬ		Cryo	(%)	(%)	(%)	(%)	(%)	(%)	(wt%)	(wt%)	(wt%)
	% Untreated			32			-9.82		0.14	-9.12	I	0.10	6.66	5.1	
	% Untreated			∞			-9.81	0.01	0.14	-8.97	0.15	0.13	93.8	5.8	
	% Untreated			8			-9.83	-0.01	0.15	-8.95	0.17	0.14	77.9	6.3	
	% Untreated			8			-9.84	-0.02	0.08	-8.86	0.26	0.11	50.4	7.3	
	% Untreated			8			-9.86	-0.04	0.11	-8.56	0.56	0.32	24.6	2.3	
10% 90%	% Untreated			11			-9.86	-0.04	0.21	-7.70	1.42	0.85	9.2	1.6	
%56 %5	% Untreated			7			-9.78	0.04	0.11	-6.78	2.34	0.63	4.2	0.7	
0% 100%	% Untreated				4		-26.01	-16.19	2.29	-6.36	2.76	1.62	0.0033*	0.0026	
	% Untreated					2	-9.82		0.02	-9.12		0.01			
%01 10%	% Untreated					3	-9.81	0.00	0.04	-9.06	90.0	0.03			
75% 25%	_					3	-9.83	-0.01	0.02	-9.09	0.04	0.07			
	% Untreated					3	-9.85	-0.03	0.03	-9.01	0.11	0.01			
	% Untreated					_	-9.86	-0.05		-8.93	0.19				
	% H ₂ O ₂ -A	0	24	4			-9.90	-0.08	0.09	-9.18	-0.06	0.12	98.8	2.7	32.7
	% H ₂ O ₂ -A	1.7	24	4			-9.90	-0.08	0.09	-8.99	0.13	0.08	95.4	3.5	53.7
		2.5	24	4			-9.88	-0.06	0.08	-8.95	0.17	0.10	96.4	3.2	38.8
		7.5	24	4			-9.87	-0.05	0.12	-8.71	0.41	0.17	67.1	3.2	62.0
		7.5	24	4			-9.88	-0.06	0.17	-7.94	1.18	0.31	32.3	3.1	50.8
10% 90%		16.9	24	4			-9.86	-0.04	0.15	-3.42	5.70	1.26	8.5	0.7	56.2
		0	72				Na	I	I	N	I	I	N E N	I	75.3
		0.1	72	4			-9.89	-0.07	0.13	-9.14	-0.02	0.08	9.96	0.6	48.8
		0.3	72	4			-9.88	-0.06	0.14	-9.03	0.09	90.0	94.3	4.1	58.6
		9.0	72	4			-9.86	-0.04	0.07	-8.99	0.13	0.08	95.7	6.9	68.5
		6.0	72	4			-9.88	-0.06	0.11	-8.84	0.28	0.08	82.6	10.5	84.9
10% 90%		1.6	72	4			9.76	90.0	0.29	-6.00	3.12	0.31	19.6	4.5	84.9
		2.9	72	4			-12.68	-2.86	1.50	-6.71	2.41	96.0	1.9	1.5	56.2
-		18.8	24		-		-40.53	-30.71	I	2.33	11.45	I	*680.0	I	87.5
		0	24	4			-9.81	0.01	0.18	-9.06	90.0	0.08	99.2	1.8	36.0
	_	9.0	24	4			-9.84	-0.02	0.09	-8.95	0.17	0.02	102.1	5.6	39.9
	_	1.3	24	4			-9.84	-0.02	0.18	-8.93	0.19	0.05	100.9	8.0	53.3
	_	2.1	24	4			-9.83	-0.01	0.12	-8.97	0.15	0.15	100.8	5.9	71.4
25% 75%	% NaOCI-A	4.3	24	4			-9.88	-0.06	0.20	-8.86	0.26	0.14	93.5	16.0	87.1
		8.9	24	4			-9.88	-0.06	0.05	-8.80	0.32	0.33	58.3	14.4	6.06
		0	48	4			-9.83	-0.01	0.12	-9.07	0.05	0.14	98.2	3.5	15.8
25% 75%		2.1	48	4			-9.89	-0.07	0.11	-9.02	0.10	0.10	99.3	5.3	87.8
10% 90%		3.8	48	4			-9.86	-0.04	0.14	-8.95	0.17	0.08	81.4	7.5	92.7
		4.8	48	4			-9.83	-0.01	0.09	-9.00	0.12	0.17	83.2	6.1	95.4
-	% NaOCI-A	7.5	24				N E	I	I	E Z	I	I	EZ	I	100.0

			ì .											
100%	%0	Plasma		_	4	-9.84	-0.02	0.10	-9.10	0.02	90.0	99.5	0.5	1.0
100%	%0	Plasma		5	4	-9.85	-0.03	0.07	-9.12	0.00	0.14	102.6	2.1	3.3
%06	10%	Plasma	2	_	4	-9.83	-0.01	0.03	-9.10	0.02	0.08	95.9	5.9	10.4
75%	25%	Plasma	5	_	4	-9.87	-0.05	0.13	-9.09	0.03	0.08	92.9	5.9	22.0
%09	%05	Plasma	15	2	4	-9.88	-0.06	0.04	-9.12	0.00	0.05	87.9	6.9	48.6
25%	75%	Plasma	45	3	4	-9.86	-0.04	0.10	-9.17	-0.05	0.07	80.8	12.7	71.4
10%	%06	Plasma	135	5	4	-9.95	-0.13	0.12	-9.29	-0.17	0.15	48.9	12.8	84.1
2%	%56	Plasma	285	8	4	-9.92	-0.10	0.05	-9.25	-0.13	0.09	39.3	8.1	8.68
%0	100%	Plasma	750	5		Z	I		R	I	I	R		89.0
Notes: Th	Votes: The bold values	ues indicate a	an offset equal to	or higher	Notes: The bold values indicate an offset equal to or higher than 0.2%. Standard deviations values equals to or higher than 0.3% are also in bold.	ard deviations v	alues equa	ls to or hig	gher than (0.3%o are	also in bo	old.		

n is the number of tubes measured by AP (Analytical Precision 2003 CF-IRMS), GB (Gas Bench coupled with a Delta plus XP CF-IRMS) or Cryo (off-line cryogenic method and dual inlet RMS) (see Materials and methods)

Nm: Not measured. The letters A and B added after H₂O₂ and NaOCI distinguish 2 batches with different conditions of H₂O₂ and NaOCI treatments (s*ee* Materials and methods). indicates μ_{veast} value for untreated yeast and yeast treated by H_2O_2 (see Discussion) external reproducibility of $\delta^{18}O_{CO2}$ measurements worsens as $\Delta^{18}O$ increases, with values up to $\pm 0.85\%$ (2 σ).

Dolomite and dolomite-yeast mixtures— H_3PO_4 digestion at $80^{\circ}C$

The $\delta^{13}C_{CO2}$ and $\delta^{18}O_{CO2}$ values for pure dolomite INYO analyzed by classical method are $-0.47\% \pm 0.01\%$ and $1.91\% \pm 0.03\%$ respectively (2 σ , n=9; Table 2). The $\delta^{13}C_{CO2}$ values for synthetic mixtures with yeast proportions $\geq 75\%$ are shifted to more negative values, with $\Delta^{13}C$ reaching -0.66% for the sample with 95% of yeast. Surprisingly $\delta^{18}O_{CO2}$ values do not show any deviation, within uncertainty, regardless of the dolomite-yeast mixture proportion (Table 2 and Fig. 2).

Pure yeast

The $\delta^{13}C_{\rm CO2}$ value of pure yeast determined by sealed tube combustion method is $-24.52\% \pm 0.12\%$ (2σ , n=3). The $\delta^{13}C_{\rm CO2}$ and $\delta^{18}O_{\rm CO2}$ values measured by CF-IRMS for CO₂ released by yeast reacted with H₃PO₄ at 25°C are $-26.01\% \pm 2.29\%$ and $-6.36\% \pm 1.62\%$, respectively (Table 1). At 80°C, the values are of $-38.39\% \pm 2.60\%$ and $-12.19\% \pm 2.48\%$, respectively (Table 2). The precisions are particularly low possibly owing to impurities and/or non-reproducibility of the yeast digestion by H₃PO₄.

Samples treated with H₂O₂

Calcite and calcite-yeast mixtures— H_3PO_4 digestion at 25°C

Pure calcite samples were partially dissolved by H₂O₂ treatment. They lost 32.7% of their mass in the H₂O₂-A protocol and 75.3% in the longer H₂O₂-B protocol (Table 1). In spite of calcite dissolution, neither $\delta^{13}C_{CO2}$ nor $\delta^{18}O_{CO2}$ were modified. Treated synthetic mixtures show both partial dissolution of calcite and incomplete organic matter removal. Calcite dissolution can be evidenced in a diagram of mass loss vs. initial yeast proportion (Fig. 3B). Samples with yeast proportion lower than 75% plot above the 1:1 line, indicating that the mass loss was higher than initial yeast proportion, and hence that calcite must have been dissolved. For 95% of yeast, only 1.9% of calcite instead of 5% remained after treatment by the H₂O₂-B protocol, demonstrating drastic calcite dissolution. Incomplete organic matter removal is evidenced only in mixtures with initial yeast proportions ≥50%, as illustrated by their calcite content lower than the theoretically expected 100% (Table 1 and Fig. 3A). Organic matter removal was improved by the longer treatment of H₂O₂-B protocol but remained incomplete for samples with initial yeast proportions \geq 75%. The $\delta^{13}C_{CO2}$ values of sample mixtures are indistinguishable from pure calcite R0, except for 95% of yeast with a Δ^{13} C value of -2.86% (Table 1 and Fig. 1A). For O isotopes, samples for which organic matter was quantitatively removed by H₂O₂ treatment have $\delta^{18}O_{CO2}$ values similar to untreated calcite R0. In contrast, samples still containing organic matter after reaction

Table 2. $\delta^{13}C_{CO2}$ and $\delta^{18}O_{CO2}$ values of pure dolomite, synthetic samples and pure yeast. $\Delta^{13}C$ (= $\delta^{13}C_{CO2\text{-sample}} - \delta^{13}C_{CO2\text{-standard}}$) and $\Delta^{18}O$ (= $\delta^{18}O_{CO2\text{-sample}}$ $-\delta^{18}O_{CO2\text{-standard}}$) correspond to offsets from the isotopic values of pure dolomite.

Sample			Yeast/sol.	Time	_	_	_	8 ¹³ C	Δ^{13} C	\pm 2 σ	$\delta^{18}O_{CO2}$	Δ ¹⁸ Ο	\pm 26	Dolomite	\pm 26	Mass loss
Dolomite	Yeast	Method	(g/L)	(E)	AP	GB	Cryo	(%)	(%)	(%)	(%)	(%)	(%)	(wt%)	(wt%)	(wt%)
100%	%0	Untreated					6	-0.47	I	0.01	1.91	I	0.03			
%06	10%	Untreated			7	7		-0.46	0.01	0.15	1.89	-0.03	0.16	92.6	11.1	
75%	25%	Untreated			7	7		-0.44	0.03	0.12	1.88	-0.03	90.0	74.4	16.7	
%05	20%	Untreated			7	7		-0.51	-0.04	0.04	1.87	-0.04	0.15	54.9	10.3	
25%	75%	Untreated			7	7		-0.57	-0.10	0.10	1.88	-0.03	0.19	25.2	8.4	
10%	%06	Untreated			7	4		-0.76	-0.29	0.19	1.91	0.00	0.21	8.6	1.6	
2%	%56	Untreated			7	4		-1.13	99.0-	0.39	1.81	-0.10	0.18	4.5	1.3	
%0	100%	Untreated				7		-38.39	-37.92	2.60	-12.19	-14.11	2.48	0.057*	0.023	
100%	%0	H_2O_2-B		72		4		-0.48	-0.01	0.23	1.93	0.01	0.14	87.0	5.8	41.4
%06	10%	H_2O_2 -B	0.1	72		4		-0.51	-0.04	0.22	1.98	0.07	0.08	92.9	6.5	43.9
75%	25%	H_2O_2 -B	0.2	72		4		-0.52	-0.05	0.13	1.94	0.03	0.07	81.0	17.2	48.5
%05	%09	H_2O_2 -B	9.0	72		4		-0.54	-0.07	0.16	1.98	0.07	0.00	85.1	10.8	50.3
25%	75%	H_2O_2-B	1.8	72		4		-0.50	-0.03	0.22	2.01	0.10	0.14	30.3	6.4	53.8
10%	%06	H_2O_2 -B	2.7	72		4		-0.80	-0.33	0.34	1.90	-0.01	0.11	15.5	1.3	51.9
2%	%56	H_2O_2 -B	2.9	72		4		-7.01	-6.54	0.93	1.56	-0.36	0.18	8.9	1.0	51.4
%0	100%	H_2O_2-B	3.0	72		-		-43.90	-43.43	I	-2.09	-4.01	I	1.5*	I	56.9
100%	%0	NaOCI-B		48		4		-0.46	0.01	0.10	1.94	0.03	0.08	91.5	10.2	6.4
%06	10%	NaOCI-B	0.1	48		4		-0.39	0.08	0.12	1.91	0.00	0.09	90.5	13.1	45.8
75%	25%	NaOCI-B	0.3	48		4		-0.44	0.03	0.12	1.86	-0.05	0.12	92.1	15.6	38.3
%05	%09	NaOCI-B	1.0	48		4		-0.42	0.05	0.19	1.85	-0.06	90.0	91.2	13.4	58.3
25%	75%	NaOCI-B	3.0	48		4		-0.43	0.04	0.17	1.86	-0.05	0.15	93.4	6.7	80.1
10%	%06	NaOCI-B	3.9	48		4		-0.46	0.01	0.19	1.83	-0.08	0.17	87.9	7.9	9.06
2%	%56	NaOCI-B	4.8	48		4		-0.50	-0.03	0.18	1.85	-0.06	0.26	77.3	3.9	93.6
%0	100%	NaOCI-B	5.0	48				R	I	I	N	I	I	Z	I	98.7
			Yeast (mg)													
100%	%0	Plasma		_	7	7		-0.47	0.00	0.11	1.89	-0.02	0.11	8.66	5.1	6.0
%06	10%	Plasma	2	-	7	7		-0.50	-0.03	0.10	1.84	-0.07	0.09	94.5	4.7	9.5
75%	25%	Plasma	5	_	7	7		-0.52	-0.05	0.03	1.87	-0.04	0.13	86.2	7.5	20.1
%05	20%	Plasma	15	2	7	7		-0.52	-0.05	0.13	1.90	-0.01	0.13	8.98	6.4	45.5
25%	75%	Plasma	45	3	7	7		-0.61	-0.14	0.04	1.80	-0.11	0.21	78.6	21.6	0.69
10%	%06	Plasma	135	2	7	7		-0.72	-0.25	0.11	1.66	-0.25	0.20	55.7	6.2	84.6
2%	%56	Plasma	285	8	7	7		-0.93	-0.46	0.19	1.46	-0.45	0.28	38.5	14.3	89.0
%0	100%	Plasma	750	5		7		-26.22	-25.75	1.58	-25.63	-27.55	1.63	0.32*	0.055	89.0
10%	%06	Plasma-bis	135	10		4		-0.62	-0.15	0.31	1.92	0.01	0.13	64.4	23.6	86.0
2%	%56	Plasma-bis	285	15		4		-0.66	-0.19	0.31	1.70	-0.21	0.56	39.8	27.1	90.4
F		7		1-1-1-		200				17 17 17	17 17 1		77.7			

Notes: The bold values indicate an offset equal to or higher than 0.2%. Standard deviations values equal to or higher than 0.3% are also in bold.

In is the number of tubes measured by AP (Analytical Precision 2003 CF-IRMS), CB (Gas Bench coupled with a Delta plus XP CF-IRMS) or Cryo (off-line cryogenic method and dual inlet IRMS) (see Materials and methods). Nm: Not measured. * *indicates μ_{yeast} value for untreated yeast and yeast treated by H_2O_2 (see Discussion). 15415856, 2021, 8, Downloaded from https://aslopubs.onlinelibrary.wiley.com/doi/10.1002/om3.10442 by Cochrane France, Wiley Online Library on [20/10/2022]. See the Terms and Conditions (https://onlinelibrary.wiley.com/erms-and-conditions) on Wiley Online Library for rules of use; OA articles are governed by the applicable Creative Commons. Licensea.

with $\rm H_2O_2$ present large deviations in $\delta^{18}O_{\rm CO2}$ up to +5.70 ‰ with a poor analytical precision (Table 1 and Fig. 1B).

Dolomite and dolomite-yeast mixtures—H₃PO₄ Digestion at 80°C

For dolomite, only the more efficient H_2O_2 -B protocol was used. For pure dolomite samples, only minor dissolution occurred (Fig. 4A; Table 2) and neither $\delta^{13}C_{CO2}$ nor $\delta^{18}O_{CO2}$ were affected by H_2O_2 treatment (as for pure calcite samples) (Fig. 2 and Table 2). For dolomite-yeast mixtures, organic matter removal was not complete for samples containing more than 50% of yeast (Fig. 4A) and dolomite was partly dissolved, as illustrated by samples plotting above the line 1:1 in Fig. 4B. The $\delta^{13}C_{CO2}$ values of sample mixtures are indistinguishable from untreated pure dolomite INYO until a yeast proportion of 90%, which shows a minor 0.33% shift. For 95% of yeast, $\delta^{13}C_{CO2}$ shifts to strongly lighter values ($\Delta^{13}C = -6.54\%$); Table 2 and Fig. 2A). A similar pattern is observed for $\delta^{18}O_{CO2}$, but with a much smaller shift at 95% only ($\Delta^{18}O = -0.36\%$); Table 2 and Fig. 2B).

Pure yeast residues—H₃PO₄ digestion at 25 and 80°C

The yeast residue (\sim 100 mg) treated by the H₂O₂-A protocol and reacted with H₃PO₄ at 25°C had $\delta^{13}C_{CO2}$ and $\delta^{18}O_{CO2}$ values of -40.53% and 2.33%, respectively (Table 1). The external precision on these measurements was not determined




Fig 1. $\delta^{13}C_{CO2}$ (**A**) and $\delta^{18}O_{CO2}$ (**B**) values of the calcite mixture samples as a function of the yeast proportion (%). The solid gray lines represent the average isotope compositions of pure calcite (R0) and the dotted lines indicate a \pm 0.2% range. The errors are comprised within symbols.

because the residue recovered after treatment was too small for more than one analysis. The yeast residue (~90 mg) treated by the $\rm H_2O_2\text{-}B$ protocol and reacted with $\rm H_3PO_4$ at 80°C had $\rm \delta^{13}C_{\rm CO2}$ and $\rm \delta^{18}O_{\rm CO2}$ values of $\rm -43.90\%$ and $\rm -2.09\%$ respectively (Table 2). In order to further investigate optimum conditions for organic matter removal, 300 mg of pure yeast were reacted with three different volumes of $\rm H_2O_2$ (50, 100, 200 ml) and for three reaction times (24, 48, and 120 h). The results showed a limited range of mass loss, between 47.1% and 55.4%, for the different conditions. However, there was a slight increase of yeast loss with increasing reaction time (Table 3) but none with increasing volume of $\rm H_2O_2$ for a given reaction time. Increasing further the reaction time (216 h for 600 mg of yeast and 50 ml of $\rm H_2O_2$) improved the mass loss but only up to 73.7% (Table 3).

Samples treated with NaOCl

Calcite and calcite-yeast mixtures— H_3PO_4 digestion at 25°C

Like for $\rm H_2O_2$, two different NaOCl protocols were tested. Pure calcite samples were partially dissolved by both treatments, losing up to 36% of their mass (Table 1 and Fig. 3B). Neither $\delta^{13}C_{\rm CO2}$ nor $\delta^{18}O_{\rm CO2}$ were modified (Table 1 and Fig. 1). For synthetic mixtures, calcite was partially dissolved by the NaOCl treatment, as shown by higher mass loss than




Fig 2. $\delta^{13}C_{CO2}$ (**A**) and $\delta^{18}O_{CO2}$ (**B**) values of the dolomite mixture samples as a function of the yeast proportion (%). The gray lines represent the average isotope compositions of pure dolomite (INYO) and the dotted lines indicate a $\pm 0.1\%$ range. The errors are comprised within symbols.




Fig 3. Efficiency of yeast removal from calcite mixture samples. (**A**) Final proportion of calcite (%) vs. initial proportion of yeast (%). As expected, untreated samples fall on the gray -1:1 line, while treated samples fall above this line, indicating yeast loss that exceeds carbonate loss. (**B**) Mass loss (%) vs. initial proportion of yeast (%). The gray 1:1 line indicates quantitative yeast loss without carbonate loss. LTA (i.e., plasma) treated samples fall on or close to this line. Samples falling above this line indicate that carbonate loss occurred in addition to yeast loss. Samples falling below this line indicate non-quantitative yeast loss.

expected from initial yeast proportion (i.e., most data plotted above the line 1:1 in Fig. 3B). Organic matter removal was incomplete. For the NaOCl-A protocol, samples with a yeast proportion \geq 75% present calcite contents lower than 100% (from 93.5 to 58.3%) (Table 1 and Fig. 3A). For the longer NaOCl-B protocol, organic matter removal was improved but remained incomplete for yeast proportions \geq 90%, with calcite contents in the residue ranging from 81.4% to 83.2%. In terms of isotope composition, the $\delta^{13}C_{CO2}$ values are indistinguishable from untreated pure calcite R0 irrespectively of the protocol or yeast content (Table 1 and Fig. 1A). $\delta^{18}O_{CO2}$ values were only slightly shifted in protocol A for yeast contents \geq 75% (with $\Delta^{18}O$ values up to +0.32%), while no shift was identified in protocol B (Table 1 and Fig. 1B).

Dolomite and dolomite–yeast mixtures— H_3PO_4 digestion at $80^{\circ}C$

Only NaOCl-B protocol was used for dolomite samples. Pure dolomite was slightly dissolved (6.4% of mass loss; Table 2 and Fig. 4B) with no modification of $\delta^{13}C_{CO2}$ and $\delta^{18}O_{CO2}$ values




Fig 4. Efficiency of yeast removal from dolomite mixture samples: (**A**) Final proportion of dolomite (%) vs. initial proportion of yeast (%). As expected, untreated samples fall on the gray -1:1 line, while treated samples fall above this line, indicating yeast loss that exceeds carbonate loss. (**B**) Mass loss (%) vs. initial proportion of yeast (%). The gray 1:1 line indicates quantitative yeast loss without carbonate loss. LTA (i.e., plasma) treated samples fall on or close to this line. Samples falling above this line indicate that carbonate loss occurred in addition to yeast loss. Samples falling below this line indicate non-quantitative yeast loss.

Table 3. Mass loss of yeast reacted with H_2O_2 and NaOCl for different conditions of time and volume of H_2O_2 or NaOCl.

H ₂ O ₂ (ml)	NaOCI (ml)	Yeast (mg)	Yeast/sol. (g/l)	Time (h)	Mass loss (wt%)
100		300	3.0	24	47.1
100		300	3.2	48	51.3
100		300	3.0	120	55.4
50		300	6.3	48	48.3
200		300	1.6	48	50.2
50		600	12.0	216	73.7
	30	300	10.0	24	92.6
	30	300	9.9	48	97.0
	30	300	10.6	120	93.8
	15	300	22.2	48	89.0
	70	300	4.4	48	98.2
	30	600	20.0	216	97.3

15415856, 2021, 8, Downloaded from https://aslopubs.onlinelibrary.wiley.com/doi/10.1002/lom3.10442 by Cochrane France, Wiley Online Library on [20/10/2022]. See the Terms and Conditions (https://onlinelibrary.wiley.com/erres-ad-conditions) on Wiley Online Library for rules of use; OA articles are governed by the applicable Certaive Commons License

Table 4. Calculated $\delta^{13}C_{CO2}$ and $\delta^{18}O_{CO2}$ values for the CO₂ produced by calcite mixture samples during H₃PO₄ digestion, assuming CO₂ originates from yeast and calcite (see text for calculation). The calculated values are compared to the measured values (also displayed in Table 1).

מומ	ר (שבר הראו	מוום כמוכוני (זכר נכאר וסו כמוכמומנוסון). דווכ כמוכמומנים	11. 11. Can	- 1	des are comb	מוכם גם מוכ וו	values are compared to the incasared values (also displayed in Table 1).	duso dispidy	מייון ומסוכיו)			
Sample			()	0,004	8 ¹³ C	Λ ¹³ C	8 ¹³ C.	Δ ¹³ C ₂₃ 2	8 ¹⁸ 0	Λ ¹⁸ Ο ₂₂₂ Ω	8 ¹⁸ O.	A ¹⁸ O _{52[23]}
Calcite	Yeast	Method	(%)	(%)	(%o)	(%o) (%o)	(%o)	(%o)	(%o)	(%o)	(%o)	(%o)
100%	%0	Untreated	100.0	0.0	-9.82	1	-9.82		-9.12	I	-9.12	
%06	10%	Untreated	93.8	6.2	-9.81	0.01	-9.82	0.00	-8.97	0.15	-9.12	0.00
75%	25%	Untreated	77.9	22.1	-9.83	-0.01	-9.82	0.00	-8.95	0.17	-9.12	0.00
%05	20%	Untreated	50.4	49.6	-9.84	-0.02	-9.82	0.00	-8.86	0.26	-9.12	0.00
25%	75%	Untreated	24.6	75.4	-9.86	-0.04	-9.82	0.00	-8.56	0.56	-9.12	0.00
10%	%06	Untreated	9.5	8.06	-9.86	-0.04	-9.83	-0.01	-7.70	1.42	-9.12	0.00
2%	%56	Untreated	4.2	95.8	-9.78	0.04	-9.83	-0.01	-6.78	2.34	-9.12	0.00
%0	100%	Untreated	0.0	100.0	-26.01	I	-26.01	1	-6.36	I	-6.36	1
100%	%0	H ₂ O ₂ -A	100.0	0.0	-9.90	-0.08	-9.82	1	-9.18	-0.06	-9.12	1
%06	10%	H ₂ O ₂ -A	95.4	4.6	-9.90	-0.08	-9.82	0.00	-8.99	0.13	-9.12	0.00
75%	25%	H ₂ O ₂ -A	96.4	3.6	-9.88	-0.06	-9.82	0.00	-8.95	0.17	-9.12	0.00
%05	20%	H ₂ O ₂ -A	67.1	32.9	-9.87	-0.05	-9.83	-0.01	-8.71	0.41	-9.11	0.01
25%	75%	H ₂ O ₂ -A	32.3	67.7	-9.88	-0.06	-9.88	-0.06	-7.94	1.18	-9.10	0.02
10%	%06	H ₂ O ₂ -A	8.5	91.5	-9.86	-0.04	-10.11	-0.29	-3.42	5.70	-9.01	0.11
%06	10%	H_2O_2 -B	9.96	3.4	-9.89	-0.07	-9.82	0.00	-9.14	-0.02	-9.12	0.00
75%	25%	H_2O_2 -B	94.3	5.7	-9.88	-0.06	-9.82	0.00	-9.03	0.09	-9.12	0.00
%05	20%	H_2O_2 -B	95.7	4.3	-9.86	-0.04	-9.82	0.00	-8.99	0.13	-9.12	0.00
25%	75%	H_2O_2 -B	82.6	17.4	-9.88	90.0-	-9.83	-0.01	-8.84	0.28	-9.12	0.00
10%	%06	H_2O_2 -B	19.6	80.4	9.76	90.0	-9.93	-0.11	-6.00	3.12	-9.08	0.04
2%	%56	H_2O_2 -B	1.9	98.1	-12.68	-2.86	-11.18	-1.36	-6.71	2.41	-8.61	0.51
%0	100%	H ₂ O ₂ -A	0.0	100.0	-40.53	l	-40.53	1	2.33	I	2.33	1

15415856, 2021, 8, Downloaded from https://aslopubs.onlinelibrary.wiley.com/doi/10.1002/lom3.10442 by Cochrane France, Wiley Online Library on [20/10/2022]. See the Terms and Conditions (https://onlinelibrary.wiley.com/erres-ad-conditions) on Wiley Online Library for rules of use; OA articles are governed by the applicable Certaive Commons License

Table 5. Calculated $\delta^{13}C_{CO2}$ and $\delta^{18}O_{CO2}$ values of the CO₂ produced by dolomite mixture samples during H₃PO₄ digestion, assuming CO₂ originates from veset and delomite (see text for calculation). The calculated values are compared to the measured values (also displayed in Table 2).

yeast and do	lomite (sa	yeast and dolomite (see text for calculation). The calculated values are compared to the measured values (also displayed in Table 2)	culation). The	e calculated	y values are c	ompared to t	the measured	values (also c	Isplayed In 18	able 2).		
Sample			() dolomito	0	8 ¹³ Cm.23	Δ ¹³ C _{m33}	8 ¹³ Cestent	A ¹³ Celent	8 ¹⁸ O _{mos}	Δ ¹⁸ O ₂₂₃₂	8 ¹⁸ O _{calcul}	A ¹⁸ O _{Calcul}
Dolomite	Yeast	Method	(wt%)	(wt%)	(%0)	(%0)	(%0)	(%0)	(%0)	(%0)	(%0)	(%0)
100%	%0	Untreated	100.0	0	-0.47	ı	-0.47	I	1.91	I	1.91	
%06	10%	Untreated	92.6	7.4	-0.46	0.01	-0.47	00.0	1.89	-0.03	1.91	0.00
75%	25%	Untreated	74.4	25.6	-0.44	0.03	-0.48	-0.01	1.88	-0.03	1.91	0.00
%05	%09	Untreated	54.9	45.1	-0.51	-0.04	-0.49	-0.02	1.87	-0.04	1.90	-0.01
25%	75%	Untreated	25.2	74.8	-0.57	-0.10	-0.53	-0.06	1.88	-0.03	1.89	-0.02
10%	%06	Untreated	9.8	90.2	-0.76	-0.29	-0.67	-0.20	1.91	0.00	1.84	-0.07
5%	%56	Untreated	4.5	95.5	-1.13	-0.66	-0.92	-0.45	1.81	-0.10	1.74	-0.17
%0	100%	Untreated	0.0	100.0	-38.39	I	-38.39	1	-12.19	I	-12.19	1
100%	%0	H_2O_2 -B	100.0	0.0	-0.48	-0.01	-0.47	1	1.93	0.01	1.91	1
%06	10%	H_2O_2 -B	92.9	7.1	-0.51	-0.04	-0.52	-0.05	1.98	0.07	1.91	0.00
75%	25%	H_2O_2 -B	81.0	19.0	-0.52	-0.05	-0.63	-0.16	1.94	0.03	1.90	-0.01
20%	%09	H_2O_2 -B	85.1	14.9	-0.54	-0.07	-0.59	-0.12	1.98	0.07	1.90	-0.01
25%	75%	H_2O_2 -B	30.3	2.69	-0.50	-0.03	-1.95	-1.48	2.01	0.10	1.77	-0.14
10%	%06	H_2O_2 -B	15.5	84.5	-0.80	-0.33	-3.82	-3.35	1.90	-0.01	1.60	-0.31
5%	%56	H_2O_2 -B	8.9	91.1	-7.01	-6.54	-6.37	-5.90	1.56	-0.36	1.37	-0.54
%0	100%	H_2O_2 -B	0.0	100.0	-43.90	I	-43.90	I	-2.09	I	-2.09	I
100%	%0	Plasma	100	0	-0.47	0.00	-0.47	I	1.89	-0.02	1.91	I
%06	10%	Plasma	94.5	5.5	-0.50	-0.03	-0.47	0.00	1.84	-0.07	1.91	-0.01
75%	25%	Plasma	86.2	13.8	-0.52	-0.05	-0.48	-0.01	1.87	-0.04	1.90	-0.01
%05	%09	Plasma	8.98	13.2	-0.52	-0.05	-0.48	-0.01	1.90	-0.01	1.90	-0.01
25%	75%	Plasma	78.6	21.4	-0.61	-0.14	-0.49	-0.02	1.80	-0.11	1.89	-0.02
10%	%06	Plasma	55.7	44.3	-0.72	-0.25	-0.54	-0.07	1.66	-0.25	1.84	-0.07
2%	%56	Plasma	38.5	61.5	-0.93	-0.46	-0.60	-0.13	1.46	-0.45	1.77	-0.14
10%	%06	Plasma-bis	64.4	35.6	-0.62	-0.15	-0.52	-0.05	1.92	0.01	1.86	-0.05
5%	%56	Plasma-bis	39.8	60.2	-0.66	-0.19	-0.59	-0.12	1.70	-0.21	1.78	-0.13
%0	100%	Plasma	0	100	-26.22	I	-26.22	I	-25.63	I	-25.63	I




Fig 5. Isotopic shifts Δ^{13} C (**A**) and Δ^{18} O (**B**) of mixture samples as a function of the external reproducibility on isotope measurements (2σ).

(Table 2 and Fig. 2). For synthetic mixtures, yeast was quantitatively removed by NaOCl, except for proportions of 90 and 95% (Table 2 and Fig. 4A) in which a small fraction of organic matter was left (12% and 23%, respectively). The NaOCl-B treatment also dissolved part of the dolomite, as evidenced by the data plotting above the line 1:1 in Fig. 4B. The $\delta^{13}C_{CO2}$ and $\delta^{18}O_{CO2}$ values of all treated mixtures were indistinguishable from untreated dolomite INYO (Table 2 and Fig. 2).

Pure yeast residues—H₃PO₄ digestion at 25°C and 80°C

The amounts of residual yeast samples treated with NaOCl were too low to be collected and prepared for isotope measurements. Nonetheless, the efficiency of organic matter removal by NaOCl has been investigated for different set of conditions, following the same procedure as for H_2O_2 . Three hundred milligrams of pure yeast was reacted with three different volumes of NaOCl (15, 30, 70 ml) and three different reaction times (24, 48, and 120 h). The results showed that organic matter removal was relatively efficient, with weight losses higher than 89% (Table 3). In contrast to H_2O_2 treatment, the weight loss of yeast increased with the volume of NaOCl for a given reaction time, while the reaction time had no effect for a given volume.

Samples treated with low-temperature oxygen-plasma ashing (LTA)

Calcite and calcite-yeast mixtures—H₃PO₄ digestion at 25°C

The mass and isotope composition of pure calcite samples were not affected by LTA treatment (Table 1 and Figs. 1 and 3B). For




Fig 6. Comparison between the isotope shifts measured for mixture samples and those calculated assuming a CO_2 contribution from yeast. The gray line represents the 1:1 line where the points should align in case of a CO_2 contribution from yeast.

synthetic mixtures, Fig. 3B illustrates no calcite loss with LTA treatment since all data plot either on or slightly below the 1:1 line. Organic matter removal was quantitative only in samples with less than 25% of yeast. For higher proportions, the calcite content ranged from 92.9% to 39.3% after LTA treatment (implying a residual organic matter content of 7.1–60.7%) (Table 1 and Fig. 3A). The $\delta^{13}C_{CO2}$ and $\delta^{18}O_{CO2}$ values for all synthetic mixtures are indistinguishable from those of untreated calcite R0, within uncertainty (Table 1 and Fig. 1).

Dolomite and dolomite-yeast mixtures— H_3PO_4 digestion at $80^{\circ}C$

The mass and isotope composition of pure dolomite samples were not affected by LTA treatment (Table 2 and Figs. 2 and 4B). For dolomite mixtures, no loss of dolomite was induced by LTA treatments since all data plot on the 1:1 line or slightly below in Fig. 4B. Similarly to calcite mixtures, organic matter removal was quantitative only in samples with less than 25% of yeast. Above this proportion, dolomite contents ranged from 86.2% to 38.5% (Table 2 and Fig. 4A) indicating the presence of residual organic matter after treatment. The $\delta^{13}C_{\rm CO2}$ and $\delta^{18}O_{\rm CO2}$ values of samples with yeast proportions $\geq 75\%$ are slightly shifted, with $\Delta^{13}C$ and $\Delta^{18}O$ values down to -0.46% and -0.45% for the sample with 95% of yeast (Table 2 and Fig. 2). The residues of samples with yeast proportions higher than 75% were thus treated again

for 5 and 7 additional hours, respectively (they are labeled "plasma-bis" in Table 2). Although organic matter was still not completely removed, the C and O isotope data were improved, with deviations lower than -0.21%0 relative to pure dolomite (Table 2).

Pure yeast residues—H₃PO₄ digestion at 25 and 80°C

After 5 h of treatment by LTA, only $\sim\!80$ mg of residue were collected from the 750 mg of pure yeast initially introduced, demonstrating that 90% of the yeast was removed. Since LTA treated mixtures reacted with H_3PO_4 at a temperature of 25°C do not present any specific isotope shift, this residue was used only for measuring the isotope composition of the CO_2 produced during an acid digestion at $80^{\circ}C$. The residue was divided in two subsamples of $\sim\!25$ and 50 mg and reacted with H_3PO_4 at $80^{\circ}C$. The average $\delta^{13}C_{CO2}$ and $\delta^{18}O_{CO2}$ values were -26.22%0 \pm 1.58%0 and -25.63%0 \pm 1.63%0, respectively (Table 2).

Discussion

Origin of the isotope shifts

In earlier studies, the main cause suspected to have induced shifts in $\delta^{13}C_{CO2}$ and $\delta^{18}O_{CO2}$ of carbonates associated with organic matter is the production of CO₂ and/or other molecular gaseous species derived from organic matter and/or impure H₃PO₄ (Bowen 1966; Epstein et al. 1951, 1953; Falster et al. 2018; Oehlerich et al. 2013; Weber et al. 1976) and their subsequent incomplete separation prior to gas introduction in the mass spectrometer. Chromium oxide or H₂O₂ was hence added to H₃PO₄ to oxidize any organic contaminants (McCrea 1950). However, these additives may also have oxidized the sample organic matter, leading to erroneous data (Sarkar et al. 1990). This was solved using H₃PO₄ of purer quality, without chromium oxide or H₂O₂. Several studies have then shown that organic matter removal was no more required for samples with low organic content (e.g., Wierzbowski 2007). However, when the weight ratio of immature organic matter vs. carbonate is high (i.e., >25%), the present results show that this is not the case, in good agreement with three previous studies (Oehlerich et al. 2013; Lebeau et al. 2014; Falster et al. 2018). The observed shifts in $\delta^{13}C_{\rm CO2}$ and $\delta^{18}O_{\rm CO2}$ (both negative) were previously attributed to a contribution of CO2 produced by organic matter during H₃PO₄ digestion (Oehlerich et al. 2013; Lebeau et al. 2014; Falster et al. 2018). Because we measured $\delta^{13}C_{CO2}$ and δ¹⁸O_{CO2} values of CO₂ produced by pure yeast reaction with H₃PO₄, we can explore this hypothesis using the following isotope mass balance equation:

$$\begin{split} \delta^{13} C_{mixture} &= \left(\omega_{yeast} \times \mu_{yeast} \times \delta^{13} C_{yeast} + \omega_{carbonate} \times \delta^{13} C_{carbonate}\right) / \\ &\left(\omega_{yeast} \times \mu_{yeast} + \omega_{carbonate}\right) \end{split} \tag{1}$$

where ω is the weight proportion (wt%) of yeast or carbonate in the sample and $\mu_{\rm yeast}$ is the weight proportion of CO₂

produced by yeast during acid digestion expressed in equivalent weight proportion of CaCO₃ in yeast: μ_{yeast} is 0.0033% \pm 0.0026% (2 σ) at 25°C (Table 1) and 0.057% \pm 0.023% (2 σ) at 80°C (Table 2). Using Eq. 1 and the isotopic values of the CO₂ produced by yeast reacted with H₃PO₄, the expected $\delta^{13}C_{CO2}$ and $\delta^{18}O_{CO2}$ values of mixture samples can be calculated and compared to the measured values (Tables 4 and 5).

For untreated calcite-yeast mixtures, the calculated and measured C and O isotope compositions of the CO₂ produced after H₃PO₄ digestion do not match (Table 4). The calculated results show that the amount of CO₂ generated by yeast reaction at 25°C is too small to induce a shift in $\delta^{13}C_{CO2}$ and $\delta^{18}O_{CO2}.$ While this is indeed the case for $\delta^{13}C_{CO2},$ it is not true for $\delta^{18}O_{CO2}$, which presents a positive shift for samples containing more than 25% of yeast (Table 4 and Fig. 6). This implies that other molecules were likely produced during yeast digestion by H_3PO_4 at 25°C. The modification of $\delta^{18}O_{CO2}$ value suggests that at least one of these molecules was not properly separated from CO₂ by the chromatographic columns used in this study (stainless column 0.3 m \times 1/8" x 2 mm, packed with Hayesep Q 60/80 mesh, Chrompack for the AP-2003 and fused silica Poraplot Q column 25 m × 0.32 mm for the Gas Bench). Moreover, the increase of $\delta^{18}O_{CO2}$ but constancy of $\delta^{13}C_{CO2}$ indicates that, in the ionization chamber, the poorly separated molecules must form molecules or radicals of m/z = 46, but not 45. Several molecules, possibly interfering on m/z from 44 to 48, have already been envisaged in previous studies: C₃H₈ (44, 45, 46), C₂H₅OH (46, 47, 48), CH₃COOH (after fragmentation in the ionization chamber: 44, 45, 46, 47); CS (44, 45, 46), BCl (45, 47), NO₂ (46, 48, 47), N₂O (44, 45, 46) (Bowen 1966; Charef and Sheppard 1984; De Groot 2011; Epstein et al. 1951; Epstein et al. 1953; Emiliani 1966; Mucciarone and Williams 1990; Weber et al. 1976). Among these, the absence of interference at m/z = 45 restricts the possible molecules to C_2H_5OH (46, 47, 48) and NO₂ (46, 48, 47). To characterize further this contamination, a few measurements of untreated calcite-yeast mixture samples (10%, 25%, 75%, and 90% of yeast) were performed using the classical method of McCrea (1950), which involves cryogenic purification of CO2 from possible contaminants in a vacuum line. In our laboratory CO2 purification is performed under vacuum at −130°C, and both C₂H₅OH and NO₂ are theoretically condensable. No significant shift (>0.2%) was observed (Table 1), which indicates that the contaminant is cryogenically separated from CO₂ under vacuum at −130°C and that C₂H₅OH and/or NO_2 might be responsible for the interference at m/z 46. Their formal identification however remains to be performed.

For untreated dolomite–yeast mixtures, the results of the calculation using Eq. 1 are compared well with the measured data (Table 5 and Fig. 6). A negative shift is found for $\delta^{13}C_{CO2}$ in samples with a proportion of yeast $\geq 75\%$ and none for $\delta^{18}O_{CO2}$ due to a smaller difference between yeast and dolomite $\delta^{18}O_{CO2}$. A contribution of CO_2 produced from the reaction of yeast with H_3PO_4 at $80^{\circ}C$ is thus consistent with $\Delta^{18}O_{CO2}$ and $\Delta^{13}C$ values measured on dolomite–yeast mixtures and

there is no need for another interfering molecule. The fact that the interference observed on m/z 46 in calcite mixtures (H₃PO₄ digestion at 25°C) is not observed in dolomite mixtures (H₃PO₄ digestion at 80°C) suggests that the contaminating molecule is either not produced, or more likely, not stable at 80°C.

To summarize, for high proportion of immature organic matter relative to carbonate, two scenarios are possible depending on the temperature of $\rm H_3PO_4$ digestion. At $25^{\circ}\rm C$, $\rm CO_2$ production from yeast reaction with $\rm H_3PO_4$ is generally negligible. However, a contaminant species contributes to the signal on mass 46 (but not 45), impacting the measurement of $\delta^{18}\rm O_{CO2}$ only. At $80^{\circ}\rm C$, the amount of $\rm CO_2$ generated by organic matter can impact both $\delta^{13}\rm C_{CO2}$ and $\delta^{18}\rm O_{CO2}$ measurements when the organic over carbonate mass ratio is high, provided that its $\delta^{13}\rm C$ and $\delta^{18}\rm O$ values are different from those of the carbonate sample.

Evaluating the need for organic matter removal

For C isotope analyses: $\delta^{13}C_{CO2}$ shifts result from the contribution of CO₂ produced by yeast during H₃PO₄ digestion. In the case of H₃PO₄ digestion at 25°C for 4 h, our results show that the production of CO₂ from yeast is too small to impact $\delta^{13}C_{CO2}$ analyses. Using Eq. 1 for a sample with 95% of yeast, a significant $\delta^{13}C_{CO2}$ shift ($\Delta^{13}C > 0.2\%$) can only be obtained if the difference between carbonate and organic matter $\delta^{13}C_{CO2}$ values reaches ~350%, which never happens in natural systems. Therefore, organic matter removal of a specific sample is not required if carbonate is dissolved at 25°C for 4 h (for instance calcite, aragonite, vaterite, hydromagnesite, or magnesite). In the case of H₃PO₄ digestion at 80°C for 2 h, a larger amount of CO2 is produced by yeast. A significant δ¹³C_{CO2} shift is produced for dolomite-yeast samples with a proportion of yeast ≥75% (Table 2 and Fig. 2A). Our previous work demonstrated that at even higher temperature (i.e., 130°C), the proportion of CO₂ produced by H₃PO₄ digestion of yeast increases so that $\delta^{13}C_{CO2}$ shifts are observed for yeast proportions as low as 25% (Lebeau et al. 2014). Hence, organic matter from organic-rich natural samples must be removed more cautiously for C isotope analysis of carbonate when acid digestion is performed at high temperature.

For O isotope analyses: $\delta^{18}O_{CO2}$ shifts are only observed for calcite samples with a proportion of yeast higher than 25% (Table 1 and Fig. 1B). These isotopic shifts result from a gaseous contaminant (possibly C_2H_5OH and/or NO_2) produced by the reaction of organic matter with phosphoric acid at 25°C. Its effect on O isotope composition is poorly reproducible and starts at relatively low proportions of organic matter. No shift in $\delta^{18}O_{CO2}$ is observed for dolomite samples, although a contribution of CO_2 issued from the reaction of yeast with H_3PO_4 at $80^{\circ}C$ is demonstrated from $\delta^{13}C_{CO2}$ results. The absence of shift likely results from a coincidence, $\delta^{18}O_{CO2}$ values of the CO_2 produced by yeast being similar to that produced by dolomite during H_3PO_4 reaction.

Our results on calcite and dolomite thus indicate that $\delta^{13}C_{CO2}$ and $\delta^{18}O_{CO2}$ may be shifted by the presence of organic matter but only when its proportions relative to carbonate are higher than 75% for δ^{13} C and 25% for δ^{18} O. The finding that below 25% of organic matter no shift is observed is compatible with most previous studies (e.g., Allison et al. 1996; McConnaughey 1989; Nagtegaal et al. 2012; Oehlerich et al. 2013; Sarkar et al. 1990; Serrano et al. 2008; Watanabe et al. 2001; Zhang et al. 2001) in which no significant shifts was observed between treated and untreated samples. Few cases of $\delta^{13}C$ and/or $\delta^{18}O$ shifts have nonetheless been reported (e.g., Allison et al. 1996; Boiseau and Juillet-Leclerc 1997; Grottoli et al. 2005; Land et al. 1975; Wierzbowski 2007). In the light of our results, a possible explanation for these shifts may be that the treatments failed at efficiently removing organic matter and instead increased the reactivity of organic matter with H₃PO₄ (as shown here in H₂O₂ treatments; Figs. 1 and 2). In those cases, the shifts might thus be due to adverse effects of the treatment on the reactivity of organic matter during carbonate digestion.

To summarize, organic matter removal is useless when the proportion of organic matter relative to carbonate is equal to or lower than 25%, even for acid digestion at 80°C. Above 25%, isotopic shifts may occur except for calcite $\delta^{13}C_{CO2}$. The reactivity of immature organic matter during acid digestion at a given temperature is probably strongly variable from one sample to another. In the case of organic matter proportions relative to carbonate higher than 25%, it is recommended to test on a few samples if organic matter removal is necessary. These tests are usually performed by evaluating the impact of organic matter removal treatment on the measured $\delta^{18}O$ and δ^{13} C values, which is time consuming. Our results show that a strong deterioration of the reproducibility is correlated with the isotopic shifts for both $\delta^{13}C_{CO2}$ and $\delta^{18}O_{CO2}$ (Tables 1 and 2; Fig. 5). A poor reproducibility can thus be used as an indicator of the need for removing organic matter from the samples.

Comparison of treatments for organic matter removal

In the absence of yeast, none of the treatments tested in the present contribution altered the isotope composition of pure carbonates, as expected from previous works (Emiliani 1966; Epstein et al. 1951, 1953; Mook, 1971; Wierzbowski 2007). For mixture samples, the efficiencies of yeast removal and reduction of the isotopic shifts induced by its presence vary from one treatment to the other. H₂O₂ treatment is not adapted to samples with large proportion of organic matter. Indeed, yeast is not quantitatively removed when its proportion is higher than 50% (Figs. 3A and 4A). This result is consistent with those of previous studies, suggesting that H₂O₂ is less efficient than NaOCl (Falster et al. 2018; Gaffey et al. 1991; Gaffey and Bronnimann 1993; Wierzbowski 2007). Moreover, our study demonstrates that after treatment, yeast is more reactive to H₃PO₄ digestion. The proportion of CO₂ produced by pure yeast increased by more than 27 times after treatment both at 25°C and 80°C ($\mu_{H2O2} = 27 \times \mu_{untreated}$) (* in Tables 1 and 2). The carbonate C and O isotope values of mixture

samples were more strongly shifted after treatment. For instance, the Δ^{13} C of dolomite with 95% of yeast is only of -0.66% without treatment and increases sharply to −6.54‰ after H₂O₂ treatment (Table 2). The isotope mass balance in Eq. 1 allows us to calculate $\delta^{13}C_{CO2}$ and $\delta^{18}O_{CO2}$ of mixture samples treated by H₂O₂ based on the quantity and isotope composition of CO₂ released by pure yeast and pure carbonates. For dolomite samples, the results of this calculation compare well with the measured data since negative shifts in $\delta^{13}C_{CO2}$ and $\delta^{18}O_{CO2}$ are obtained for samples containing more than 75% of yeast (Table 5 and Fig. 6). The calculated shifts are larger than the measured values, probably reflecting inaccurate determination of the isotope compositions and μ of yeast treated by H_2O_2 (only one measurement). Nonetheless they are consistent with a contribution of CO₂ produced by yeast treated with H₂O₂. As for untreated samples, this explanation does not account for the behavior of calcite samples. Although at 25°C, the yeast treated by H₂O₂ is also more reactive than untreated yeast, the shifts measured in $\delta^{18}O_{CO2}$ are not predicted by the calculation from Eq. 1, suggesting the presence of others molecules (Table 4 and Fig. 6B). Finally another drawback of H₂O₂ treatment is the significant carbonate dissolution, as previously documented (Falster et al. 2018; Gaffey and Bronnimann 1993; Nagtegaal et al. 2012; Pingitore et al. 1993; Wierzbowski 2007). This is illustrated in this study by a higher mass loss after treatment than the initial mass proportion of yeast (Figs. 3B and 4B). This may be an issue for natural samples with isotopically heterogeneous carbonate grains. In that case, carbonate dissolution may be prevented by using an alkaline H₂O₂ solution (Fallet et al. 2009; Falster et al. 2018).

In the present contribution, the NaOCl treatment did not remove organic matter quantitatively (as also noted by Charef and Sheppard 1984) but was much more efficient than H₂O₂ and LTA (Figs. 3A and 4A). In the worst case, less than 25% of the initial amount of yeast remained after treatment and most importantly, $\delta^{13}C_{CO2}$ and $\delta^{18}O_{CO2}$ values of treated mixture samples were indistinguishable from those of untreated pure carbonates, i.e., calcite RO and dolomite INYO. Although previous studies reported no appreciable dissolution of calcium carbonate by NaOCl treatment (Gaffey and Bronnimann 1993; Nagtegaal et al. 2012; Pingitore et al. 1993), our experimental results clearly demonstrate carbonate dissolution, yet to a smaller extent than H₂O₂ (Figs. 3B and 4B). It is thus not recommended to use NaOCl treatment for small and/or isotopically heterogeneous samples, unless the pH of the solution can be made more alkaline (e.g., Nagtegaal et al. 2012).

The low-temperature oxygen plasma treatment removes less organic matter than NaOCl but more than H_2O_2 (Figs. 3A and 4A). The yeast reactivity to H_3PO_4 digestion is increased by LTA although less than by H_2O_2 , with a production of CO_2 five times more important than untreated yeast ($\mu_{plasma} = 5 \times \mu_{untreated}$) (* in Table 2). However, organic matter removal by LTA is efficient enough so that $\delta^{13}C_{CO2}$ and $\delta^{18}O_{CO2}$ of the mixture samples are indistinguishable from those of untreated pure carbonates R0 and INYO. This is also supported by the good agreement in

isotope mass balance between predicted and measured CO_2 in yeast–dolomite mixtures (Table 5 and Fig. 6). A significant exception is observed for dolomite samples with 95% of yeast that present a $\Delta^{18}O$ of -0.45% after 8 h of LTA treatment, and -0.21% after 15 h of treatment (Table 5). The main advantage of plasma ashing is that it occurs in dry conditions, hence preventing carbonate dissolution and allowing perfect preservation of carbonates: only 1% of mass loss occurs for pure carbonate samples after LTA treatment while wet treatments induce between 6% and 40% of mass loss (Tables 1 and 2; Figs. 3 and 4). This method thus offers a possibility to remove organic matter from carbonate-poor samples while allowing the quantification of the carbonate content in the initial sample, provided that the mass loss induced by LTA treatment is quantified.

Conclusion

In the present work, we compared the three classical treatment used for organic matter removal. Our aim was to evaluate both their necessity and usefulness for calcite or dolomite isotope measurements using the CF-IRMS method, which has nearly completely replaced dual inlet IRMS methods in stable isotope laboratories worldwide. The need for organic matter removal in carbonate samples was evaluated for a wide range of immature organic matter proportions, with a focus on the extremely high proportions under-investigated so far.

Our results show that when using CF-IRMS, no treatment is necessary for organic matter over carbonate ratios lower than 50%, i.e., for most natural samples. For higher ratios, we recommend to test if organic matter removal is necessary especially when a poor reproducibility is observed on $\delta^{13}C_{\rm CO2}$ and $\delta^{18}O_{\rm CO2}$ measurements of untreated samples. In the present experiments that use yeast as a model for immature organic matter, the results show that for untreated calcite mixtures, the $\delta^{13}C_{\rm CO2}$ values were never affected while $\delta^{18}O_{\rm CO2}$ can be shifted (up to +2‰) for yeast proportions higher than 50%. For dolomite, $\delta^{13}C_{\rm CO2}$ values were slightly shifted (down to -0.6‰) only for yeast proportions higher than 75%. $\delta^{18}O_{\rm CO2}$ values were not affected because of a similarity in the $\delta^{18}O_{\rm CO2}$ values of dolomite and yeast but this may not always be the case.

Our comparison of the three different treatments shows that for samples extremely rich in organic matter, NaOCl is the most efficient procedure, while $\rm H_2O_2$ is the least efficient. In terms of carbonate preservation, $\rm H_2O_2$ showed a strong dissolution of carbonates, NaOCl a mild dissolution, while LTA preserved them perfectly. Finally, both NaOCl and LTA treatments provide precise and accurate isotope results on carbonates while $\rm H_2O_2$ treatments strongly increased the reactivity of the remaining organic matter toward $\rm H_3PO_4$, therefore strongly worsening both C and O isotopic shifts.

Our results also provide constraints on the mechanisms responsible for isotope offsets linked to high proportions of organic matter relative to carbonates. During acid digestion at 25°C, yeast generated a negligible amount of CO₂ and more

importantly a volatile contaminant. This contaminant was not efficiently separated from CO_2 by chromatography and interfered on the mass 46~(m/z) shifting only $\delta^{18}O_{CO2}$ values, but was separated by cryogenic purification at $-130^{\circ}C$ under vacuum. Future work should focus on identifying this contaminant and improving its chromatographic separation from CO_2 . The use of an appropriate column should be considered once it has been identified. For example, a column Molesieve 5A can separate CO_2 from NO_2 . Finally, when carbonates are digested at $80^{\circ}C$, immature organic matter does not release this contaminant but generate much more CO_2 than at $25^{\circ}C$, which is added to the CO_2 produced by carbonate dissolution. This may be a particular concern for methods based on calcite digestion at temperature of $70^{\circ}C$ or higher.

References

- Ader, M., and M. Javoy. 1998. Early diagenesis in a sulphate-reducing environment: an isotopic study in the Lower Jurassic of the Paris Basin. Comptes Rendus de l'Academie des Sciences Series IIA Earth and Planetary Science. **327**: 803–809.
- Allison, N., A. W. Tudhope, and A. E. Fallick. 1996. Factors influencing the stable carbon and oxygen isotopic composition of *Porites lutea* coral skeletons from Phuket, South Thailand. Coral Reefs **15**: 43–57. doi:10.1007/BF01626076
- Boiseau, M., and A. Juillet-Leclerc. 1997. H₂O₂ treatment of recent coral aragonite: Oxygen and carbon isotopic implications. Chem. Geol. **143**: 171–180. doi:10.1016/s0009-2541 (97)00112-5
- Bowen, R. 1966. Paleotemperature analysis. Amsterdam: Elsevier. Bowen, R. 1991. Isotopes and climates. Barking: Elsevier Science.
- Charef, A., and S. M. F. Sheppard. 1984. Carbon and oxygen isotope analysis of calcite or dolomite associated with organic matter. Isotope Geosci **2**: 325–333. doi:10.1016/0009-2541(84)90175-x
- Conan, S. M. H., and G.-J. A. Brummer. 2000. Fluxes of planktic foraminifera in response to monsoonal upwelling on the Somalia Basin margin. Deep Sea Res II **47**: 2207–2227. doi:10.1016/S0967-0645(00)00022-9
- De Groot, P. Amsterdam: Elsevier Science; 2011. Chapter 4–5, Carbon (13 C/ 12 C) and oxygen (18 O/ 16 O) isotopes in carbonate rock and minerals, p. 273–322. *In* Handbook of stable isotope analytical techniques, v. **II**.
- D'Eugenio, M., and G. Leone. 1989. Influenza della metodologia di preparazione sull'analisi isotopica di carbonati biogenici. Quad. Mus. Stor. Nat. Livorno **10**: 1–8.
- D'Hondt, S., and M. Lindinger. 1994. A stable isotopic record of the Maastrichtian Ocean-climate system: South Atlantic DSDP site 528. Palaeogeogr. Palaeoclimatol. Palaeoecol. **112**: 363–378. doi:10.1016/0031-0182(94)90081-7
- Durazzi, J. T. 1977. Stable isotope in ostracod shell: A preliminary study. Geochim. Cosmochim. Acta **41**: 1168–1170. doi:10.1016/0016-7037(77)90113-2

- Emiliani, C. 1966. Paleotemperature analysis of Caribbean cores P6304-8 and P6304-9 and a generalized temperature curve for the last 425,000 years. J. Geol. **74**: 109–126.
- Epstein, S., R. Buchsbaum, H. A. Lowenstam, and H. C. Urey. 1951. Carbonate-water isotopic temperature scale. Geol. Soc. Am. Bull. **62**: 417. doi:10.1130/0016-7606(1951)62 [417:cits]2.0.co:2
- Epstein, S., R. Buchsbaum, H. A. Lowenstam, and H. C. Urey. 1953. Revised carbonate-water isotopic temperature scale. Geol. Soc. Am. Bull. **64**: 1316–1326. doi:10.1130/0016-7606(1953)64[1315:rcits]2.0.co;2
- Erez, J., and S. Honjo. 1981. Comparison of isotopic composition of planktonic foraminifera in plankton tows, sediment traps and sediments. Palaeogeogr. Palaeoclimatol. Palaeoecol. **33**: 129–156. doi:10.1016/0031-0182(81)90035-3
- Fallet, U., W. Boer, C. Van Assen, M. Greaves, and G.-J. A. Brummer. 2009. A novel application of wet oxidation to retrieve carbonates from large organic-rich samples for ocean climate research. Geochem. Geophys. Geosyst. **10**: Q08004. doi:10.1029/2009GC002573
- Falster, G., S. Delean, and J. Tyler. 2018. Hydrogen peroxide treatment of natural lake sediment prior to carbon and oxygen stable isotope analysis of calcium carbonate. Geochem. Geophys. Geosyst. **19**: 3583–3595. doi:10.1029/2018gc007575
- Feldmeijer, W., B. Metcalfe, P. Scussolini, and K. Arthur. 2013. The effect of chemical pretreatment of sediment upon foraminiferal-based proxies. Geochem. Geophys. Geosyst. **14**: 3996–4014. doi:10.1002/ggge.20233
- Gaffey, S. J., and C. E. Bronnimann. 1993. Effects of bleaching on organic and mineral phases in biogenic carbonates. J. Sediment. Petrol. **63**: 752–754. doi:10.1306/d4267be0-2b26-11d7-8648000102c1865d
- Gaffey, S. J., J. J. Kolak, and C. E. Bronnimann. 1991. Effects of drying, heating, annealing, and roasting on carbonate skeletal material with geochemical and diagenetic implications. Geochim. Cosmoschim. Acta **55**: 1627–1640. doi:10. 1016/0016-7037(91)90134-q
- Goreau, T. J. 1977. Coral skeletal chemistry: Physiological and environmental regulation of stable isotopes and trace metals in *Montastrea annularis*. Proc. R. Soc. Lond. B **196**: 291–315. doi:10.1098/rspb.1977.0042
- Grauel, A. L., T. W. Schmid, B. Hu, C. Bergami, L. Capotondi, L. Zhou, and S. M. Bernasconi. 2013. Calibration and application of the "clumped isotope" thermometer to foraminifera for high-resolution climate reconstructions. Geochim. Cosmochim. Acta **108**: 125–140. doi:10.1016/j.gca.2012.12.049
- Grottoli, A. G., L. J. Rodrigues, K. A. Matthews, J. E. Palardy, and O. T. Gibb. 2005. Pre-treatment effects on coral skeletal delta C-13 and delta O-18. Chem. Geol. **221**: 225–242. doi: 10.1016/j.chemgeo.2005.05.004
- Guiguer, K. R. R. A., R. Drimmie, and M. Power. 2003. Validating methods for measuring $\delta^{18}O$ and $\delta^{13}C$ in otoliths from

- freshwater fish. Rapid Commun. Mass Spectrom. **17**: 463–471. doi:10.1002/rcm.935
- Hendry, J. P., and R. M. Kalin. 1997. Are oxygen and carbon isotopes of mollusc shells reliable palaeosalinity indicators in marginal marine environments? A case study from the Middle Jurassic of England. J. Geol. Soc. London **154**: 321–333. doi:10. 1144/gsigs.154.2.0321
- Irwin, H., C. Curtis, and M. Coleman. 1977. Isotopic evidence for source of diagenetic carbonates formed during burial of organic-rich sediments. Nature **269**: 209–213. doi:10.1038/269209a0
- Katz, A., M. Bonifacie, M. Hermoso, P. Cartigny, and D. Calmels. 2017. Laboratory-grown coccoliths exhibit no vital effect in clumped isotope (Δ47) composition on a range of geologically relevant temperatures. Geochim. Cosmochim. Acta 208: 335–353. doi:10.1016/j.gca.2017. 02.025
- Keatings, K. W., J. A. Holmes, and T. H. E. Heaton. 2006. Effects of pre-treatment on ostracod valve chemistry. Chem. Geol. **235**: 250–261. doi:10.1016/j.chemgeo.2006. 07.003
- Land, L. S., J. C. Lang, and D. J. Barnes. 1975. Extension rate: A primary control on the isotopic composition of West Indian (Jamaican) scleractinian reef coral skeletons. Mar. Biol. **33**: 221–233. doi:10.1007/bf00390926
- Lebeau, O., V. Busigny, C. Chaduteau, and M. Ader. 2014. Organic matter removal for the analysis of carbon and oxygen isotope compositions of siderite. Chem. Geol. **372**: 54–61. doi:10.1016/j.chemgeo.2014.02.020
- Leone, G., F. Bonadonna and G. Zanchetta. 2000. Stable isotope record in mollusca and pedogenic carbonate from Late Pliocene soils of Central Italy. Palaeogeography, Palaeoclimatology, Palaeoecology. **163**: 115–131. doi:10.1016/S0031-0182(00)00148-6
- Loncaric, N., J. van Iperen, D. Kroon, and G.-J. A. Brummer. 2007. Seasonal export and sediment preservation of diatomaceous, foraminiferal and organic matter mass fluxes in a trophic gradient across the SE Atlantic. Prog. Oceanogr. **73**: 27–59. doi:10.1016/j.pocean.2006.10.008
- McConnaughey, T. 1989. ¹³C and ¹⁸O isotopic disequilibrium in biological carbonates: I Patterns. Geochim. Cosmochim. Acta **53**: 151–162. doi:10.1016/0016-7037(89)90282-2
- McCrea, J. M. 1950. On the isotopic chemistry of carbonates and a paleotemperature scale. J. Chem. Phys. **18**: 849–857.
- Mook, W. G. 1971. Paleotemperatures and chlorinities from stable carbon and oxygen isotopes in shell carbonate. Palaeogeography, Palaeoclimatology, Palaeoecology. **9**: 245–263. doi:10.1016/0031-0182(71)90002-2
- Mucciarone, D. A., and D. F. Williams. 1990. Stable isotope analyses of carbonate complicated by nitrogen oxide contamination: A Delaware basin example. J. Sediment. Petrol. **60**: 608–614. doi:10.1306/212F91FB-2B24-11D7-8648000102C1865D

- Nagtegaal, R., C. A. Grove, S. Kasper, J. Zinke, W. Boer, and G. J. A. Brummer. 2012. Spectral luminescence and geochemistry of coral aragonite: Effects of whole-core treatment. Chem. Geol. **318–319**: 6–15. doi:10.1016/j. chemgeo.2012.05.006
- Oehlerich, M., M. Baumer, A. Lücke, and C. Mayr. 2013. Effects of organic matter on carbonate stable isotope ratios (δ^{13} C, δ^{18} O values)—implications for analyses of bulk sediments. Rapid Commun. Mass Spectrom. **27**: 707–712. doi: 10.1002/rcm.6492
- Peral, M., and others. 2018. Updated calibration of the clumped isotope thermometer in planktonic and benthic foraminifera. Geochim. Cosmochim. Acta **239**: 1–16. doi: 10.1016/j.gca.2018.07.016
- Pingitore, N. E., S. B. Fretzdorf, B. P. Seitz, L. Y. Estrada, P. M. Borrego, G. M. Crawford, and K. Love. 1993. Dissolution kinetics of CaCO₃ in common laboratory solvents. J. Sediment. Petrol. **63**: 641–645. doi:10.1306/D4267B9A-2B26-11D7-8648000102C1865D
- Sarkar, A., R. Ramesh, and S. K. Bhattacharya. 1990. Effect of sample pretreatment and size fraction on the δ^{18} O and δ^{13} C values of foraminifera in Arabian Sea sediments. Terra Nova **2**: 488–493. doi:10.1111/j.1365-3121.1990.tb00107.x
- Serrano, O., L. Serrano, and M. G. Mateo. 2008. Effects of sample pre-treatment on the δ^{13} C and δ^{18} O values of living benthic foraminifera. Chem. Geol. **257**: 218–220. doi:10. 1016/j.chemgeo.2008.09.013
- Stevenson, E. I., and others. 2014. Controls on stable strontium isotope fractionation in coccolithophores with implications for the marine Sr cycle. Geochim. Cosmochim. Acta **128**: 225–235. doi:10.1016/j.gca.2013.11.043
- Swart, P., and M. Coleman. 1980. Isotopic data for scleractinian corals explain their palaeotemperature uncertainties. Nature **283**: 557–559. doi:10.1038/283557a0
- Tripati, A. K., R. A. Eagle, N. Thiagarajan, A. C. Gagnon, H. Bauch, P. R. Halloran, and J. M. Eiler. 2010. 13C-18O isotope signatures and "clumped isotope" thermometry in foraminifera and coccoliths. Geochim. Cosmochim. Acta **74**: 5697–5717. doi:10.1016/j.gca.2010.07.006
- Watanabe, T., M. Minagawa, T. Oba, and A. Winter. 2001. Pretreatment of coral aragonite for Mg and Sr analysis: Implications for coral thermometers. Geochem. J. **35**: 265–269. doi:10.2343/geochemj.35.265
- Weber, J. N., P. Deines, P. H. Weber, and P. A. Baker. 1976. Depth related changes in ¹³C/¹²C ratio of skeletal carbonate deposited by Caribbean reef-frame building coral Montastrea annularis: Further implications of a model for stable isotope fractionation by scleractinian corals. Geochim. Cosmochim. Acta **40**: 31–39. doi:10.1016/0016-7037(76)90191-5
- Wierzbowski, H. 2007. Effects of pre-treatments and organic matter on oxygen and carbon isotope analyses of skeletal and inorganic calcium carbonate. Int. J. Mass Spectrom. **268**: 16–29. doi:10.1016/j.ijms.2007.08.002

Carbonate isotope analyses of organic-rich sediments

Zhang, C. L., J. Horita, D. R. Cole, J. Z. Zhou, D. R. Lovley, and T. J. Phelps. 2001. Temperature dependent oxygen and carbon isotope fractionations of biogenic siderite. Geochim. Cosmochim. Acta **65**: 2257–2271. doi:10.1016/S0016-7037 (01)00596-8.

ANR-11-IDEX-0005-02), and Institut Universitaire de France funding to V.B. (IUF#2017-2021).

Submitted 28 August 2020 Revised 30 May 2021 Accepted 31 May 2021

Acknowledgments

Associate editor: Dr. Ben Surridge

This work was supported by the ANR DZIANI (ANR-13-BS06-0001), the UnivEarths Labex program at Sorbonne Paris Cité (ANR-10-LABX-0023 and